



## ANNEX

### EU TYPE EXAMINATION CERTIFICATE Nr. 852/2020/0883

#### 1. Applicant

Cambridge Environmental Solutions Ltd  
1001-2 Wilson House 19 Wyndham Street  
Central Hong-Kong  
Hong Kong

#### 2. Description

The article Cambridge mask PRO fulfils the requirements of EN 149:2001+A1:2009 as type FFP2 R D

#### 3. Materials and accessories

##### **Mask**

- Cambridge Mask PRO

## 4. Technical documentation

### Summary test results

EN 149:2001+A1:2009

Mask

**Cambridge Mask PRO**

Method	Description	Result	Class
§ 8.2 EN 149: Visual Inspection	Visual inspection	PASS	Pass
§ 8.2 EN 149: Visual Inspection	Material	PASS	Pass
§ 8.4 EN 149: Practical Performance	Practical Performance	PASS	Pass
§ 7.9.1 EN 149	Total inward leakage	PASS	FFP2
§ 7.9.2 EN 149 NaCl EN 13274-4	Penetration of filter material	PASS	FFP2
§ 7.9.2 EN 149 Paraffine EN 13274-4	Penetration of filter material	PASS	FFP2
§ 8.4 & 8.5 EN 149	Compatibility with skin	PASS	FFP2
§ 8.6 EN 149	Flammability	PASS	Pass
§ 8.7 EN 149	Carbon dioxide content of the inhalation air	PASS	Pass
§ 8.4 & 8.5 EN 149	Head Harness	PASS	FFP2
§ 8.4 EN 149	Field of vision	PASS	Pass
§ 7.15 EN 149	Exhalation valve	PASS	Pass
§ 8.9 EN 149	Breathing resistance	PASS	FFP2
§ 7.17 & 8.10 EN 149	Clogging test	PASS	Pass
§ 7.17 & 8.11 EN 149	Demountable parts	NA	NA

### Description/Picture of article

Article

**Cambridge Mask PRO**



The above picture is a general picture of the article. Possible variations of the above article can be present in the technical file.

Note :

Any modification in material, design, or other technical features must be brought to the attention of the Notified Body.